

For pendant drop experiments the air flow from ventilation may be a problem. This shield put on top of the sample stage should help.

Be careful with the placement, because there is not extra space for the disposable tip.

NOTE: there is not enough space for a microscope slide when this box is on top of the stage.

